


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 0617 Accredited to ISO/IEC 17025:2005	LTE Scientific Ltd	
	Issue No: 012 Issue date: 25 November 2016	
	Greenbridge Lane Greenfield Oldham Lancashire OL3 7EN	Contact: Kim Cooper Tel: +44 (0)1457 876221 Fax: +44 (0)1457 870131 E-Mail: k.cooper@lte-scientific.co.uk Website: www.lte-scientific.co.uk
Calibration performed by the Organisations at the locations specified below		

Locations covered by the organisation and their relevant activities

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Customer Premises, eg, Hospitals, Laboratories, and Manufacturing Plants (*including those of LTE)	Temperature (Sterilizers etc, and Thermal Products) Pressure Time	S



0617

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

LTE Scientific Ltd

Issue No: 012 Issue date: 25 November 2016

Calibration performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks	Location Code
<p>TEMPERATURE</p> <p>Temperature controlled autoclaves, media preparators, incubators, sterilizers, ovens, environmental cabinets, and liquid baths (inclusive of associated indicators, controllers and recorders, all with sensors, within the specified parameters and ranges).</p>	<p>0 °C to 140 °C 140 °C to 250 °C</p>	<p>0.63°C 1.2 °C</p>	<p>A list of individual approved engineers is held by the laboratory and by UKAS.</p> <p>Single and multipoint monitoring probes. Time dependent temperature profiling.</p>	S
<p>TIME INTERVAL</p> <p>Timers</p>	<p>30 s to 3600 s</p>	<p>4.0 s</p>		S
<p>PRESSURE</p> <p>Gas pressure (gauge)</p>	<p>- 0.9 kPa to 300 kPa</p>	<p>6 kPa</p>		S

END